







90DR45D

Parameters	Specifications
Display	4½ Digit Display
Resistance Range	0 ~ 211.111110GΩ
10 Adjustable Potentiometric	x10 <sup>2</sup> , x10 <sup>3</sup> , x10 <sup>4</sup> , x10 <sup>5</sup> , x10 <sup>6</sup> ,
Resistance	x10 <sup>7</sup> x10 <sup>8</sup> , x10 <sup>9</sup> , x10 <sup>10</sup> , x10 <sup>11</sup>
Accuracy	±0.2%, ±0.2%, ±0.2%,
	±0.2%, ±0.2%, ±0.5%, ±1%,
	±2%, ±5%, ±10%
	(respectively)
Current / Voltage Value of	50mA, 20mA, 10mA, 2mA,
Each Resistance	1000V, 2500V, 5000V, 5000V,
	5000V (respectively)
Voltage Range	0 ~ 1999.9V ~ 5000V
	(RMS / V <sub>PP</sub> )
Accuracy	±(1% rdg +2digits)
Input Impedance	10GΩ ± 10%
Insulation Resistance between	1ΤΩ
Circuit and Housing	
Maximum With Stand Voltage	10000V AC(RMS) For
Between Circuit and Housing	<60 Seconds
Working Temperature and	20 ± 5ºC, ≤ 65% RH
Humidity	
Storage Temperature and	-10 ~ 55ºC, ≤ 80% RH
Humidity	
Dimension	520 x 285 x 180 mm (approx.)
Power	9V Battery
Weight	5.3Kg (approx.)
Applications	• Checking of HV Insulation
	Testers
	Research Laboratories
	• Factories
	<ul> <li>School &amp; Institute</li> </ul>
	Accreditation Laboratories

Parameters	Specifications
Resistance Range	0.01MΩ ~ 5GΩ
17 Selectable Fixed	0.01MΩ, 0.05MΩ, 0.1MΩ,
Resistance Values	0.2ΜΩ, 0.5ΜΩ, 1ΜΩ, 2ΜΩ,
	5ΜΩ, 10ΜΩ, 20ΜΩ, 50ΜΩ,
	100MΩ, 200MΩ, 500MΩ,
	1kMΩ, 2kMΩ, 5kMΩ
Accuracy	±0.2%, ±0.2%, ±0.2%, ±0.2%,
	±0.2%, ±0.2%, ±0.2%, ±0.2%,
	±0.5%, ±0.5%, ±0.5%, ±1%,
	±1%, ±1%, ±2%, ±2%, ±2%
	(respectively)
Voltage Range (DC)	100V, 200V, 300V, 400V, 500V,
	1kV, 1kV, 1kV, 1kV, 2.5kV,
	2.5kV, 2.5kV, 2.5kV, 2.5kV,
	2.5kV, 2.5kV, 2.5kV (respectively)
Insulation Resistance between	>500GΩ
Circuit and Housing	
Maximum With Stand Voltage	2000V AC For <60 Seconds
Between Circuit and Housing	
Working Temperature	20 ± 5°C (Typical)
	20 ± 5℃ (Reference)
Storage Temperature	-10 ~ 40ºC, ≤ 75% RH
Humidity	25% ~ 75% (Typical)
	30% ~ 60% (Reference)
Dimension	310 x 240 x 160mm (approx.)
Weight	3.6Kg (approx.)
Applications	<ul> <li>Checking of HV Insulation</li> </ul>
	Testers
	<ul> <li>Research Laboratories</li> </ul>
	<ul> <li>Factories</li> </ul>
	<ul> <li>School &amp; Institute</li> </ul>
	<ul> <li>Accreditation Laboratories</li> </ul>